

LEAN-MEATER INSTRUCTIONS

ADDENDUM

CLARIFICATIONS

LAYERS SWITCH AND PROBING

For both positions of the Layers Switch, after the Read light is lit, probe a little longer to try to get a deeper reading. This will produce the correct reading in cases where your first reading resulted from an echo from irregularities in the fat.

When the Layers switch is in position “2”, the additional probing may produce a deeper reading for the 2nd layer, but it will never indicate depth to the bottom of the 3rd layer.

When the Layers switch is in position “3”, the Read light will not light until the instrument believes that it has detected the bottom of the 3rd layer.

NUMBER THAT DISPLAYS WHEN INSTRUMENT IS TURNED ON

Most Lean-Meaters will display a number when turned on. The number is most often “77”, but can be anything and may change from time to time. The number is meaningless and should be disregarded. We could have suppressed this display, but doing so would have meant an additional cost and would have provided no benefit to the customer.